

PD MiniTrap G-25

Product Booklet

Code: 28922529

PD MiniTrap G-25 contains

- 50 prepacked disposable PD MiniTrap™ columns containing 2.1 mL of Sephadex™ G-25 resin
- 4 adapters
- Instructions for use

Purpose

PD MiniTrap G-25 columns are prepacked and designed for rapid, convenient sample clean-up of proteins and other large biomolecules (>5000 Mr).

PD MiniTrap G-25 columns can be used in a wide range of applications such as desalting, buffer exchange and removal of low-molecular weight compounds



Table of contents

1. Principle	2
2. Advice on handling	4
3. Safety precautions	6
4. Column assembly	6
5. Gravity protocol	7
6. Spin protocol	8
7. Recovery and desalting capacity	9
8. Column characteristics	10
9. Ordering information	11

1 Principle

PD MiniTrap G-25 columns contain *Sephadex G-25 resin*, which allows rapid *group separation* of high molecular weight substances from low molecular weight substances.

PD MiniTrap G-25 columns are used for desalting, buffer exchange and sample clean up. Small molecules like salt, free labels and other impurities are efficiently separated from the high molecular weight substances of interest.

The chromatography technique is gel filtration and molecules are separated on the basis of differences in size

- Molecules larger than the largest pores in the Sephadex matrix are excluded from the matrix and are eluted first, in or just after the *void volume*. The void volume is the column volume outside the Sephadex matrix.
- Molecules smaller than the largest pores in the Sephadex matrix will penetrate the pores to varying extent. They have a larger accessible column volume than the large molecules and therefore they elute after the large molecules just before one *total column volume* of buffer has passed through the column.

Group separation can be made using two different protocols, gravity protocol and spin protocol.

GE provides an assortment of sample clean-up products. The different formats available are summarized in Table 1.

Table 1. Product overview

Clean Up product	Exclusion limit, M_r	Bed volume	Sample volume gravity protocol ¹	Sample volume spin protocol ¹
PD SpinTrap™ G-25	5000	0.5 mL	-	100 to 180 µL
PD MultiTrap™ G-25	5000	0.5 mL	-	70 to 130 µL
PD MiniTrap™ G-25	5000	2.1 mL	0.1 to 0.5 mL	0.2 to 0.5 mL
PD MidiTrap™ G-25	5000	3.5 mL	0.5 to 1.0 mL	0.75 to 1.0 mL
PD-10 Desalting Columns	5000	8.3 mL	1.0 to 2.5 mL	1.75 to 2.5 mL
PD MiniTrap G-10	700	2.1 mL	0.1 to 0.3 mL	-
PD MidiTrap G-10	700	5.3 mL	0.4 to 1.0 mL	-

¹ Recommended sample volumes.

2. Advice on handling

Protocol selection

The separation can be made using two different protocols, gravity protocol or spin protocol, see Table 2 for protocol overview.

Gravity protocol

The liquid passes through the column by gravity force.

- There is a slightly higher recovery and desalting capacity using gravity protocol compared to when using the spin protocol.
- The applied sample is diluted.

Spin protocol

Additional gravity force is added by spinning the column in a centrifuge for some protocol steps

- There is no dilution of the sample.

Table 2. Protocol overview

Protocol	Sample volume	Elution buffer	Dilution factor	Desalting capacity
Gravity	0.1–0.5 mL	1.0 mL	2 times ¹	>98%
Spin	0.2–0.5 mL	None	None	>90%

¹ Two times dilution valid if 0.5 mL sample volume is used.

Recovery

The recovery of applied amount sample is dependent on type of protein or other biomolecule. Typically the recovery is in the range 70–90%. An increase in sample concentration can improve recovery.

Equilibration

- It is critical to equilibrate the column since UV absorbing stabilizers are used in column packing.
- Equilibration is most conveniently made by gravity also when using the spin protocol.

Sample Application

- The MiniTrap column is intended for sample volumes up to 0.5 mL.
- For sample volumes less than 0.5 mL, allow the sample to enter the packed bed completely and then add equilibration buffer (stacker volume) so that the total volume of sample and buffer added equals 0.5 mL.
- Allow the sample to enter the packed bed completely before any addition of buffer for elution.

Elution

- For higher recoveries, use stacker volumes (spin protocol).

Centrifugation

- For better result if using a fixed angle rotor centrifuge: Put the columns in same direction in all centrifugation steps (spin protocol).

3 Safety precautions

Always use normal personal protection devices like gloves and safety glasses when handling PD MiniTrap G-25 columns.



WARNING: The column storage solution, 0.15% Kathon™ CG/ICP biocide, is potentially allergenic. Use gloves when discarding the storage solution.

4 Column assembly

For use in the spin protocol the PD MiniTrap G-25 column should be assembled with adapter and collection tube as shown in Fig 1.

Note: This column assembly may also be used for convenient handling of columns when using the gravity protocol.




Fig 1. Assembly of column, adapter and collection tube.

5 Gravity protocol



1 PD MiniTrap G-25 preparation


- Remove the top cap and pour off the column storage solution.
- Remove the bottom cap.



2 Column equilibration

- Fill up the column with equilibration buffer and allow the equilibration buffer to enter the packed bed completely.
- Repeat twice.
- Discard the flow-through.

Note: About 8 mL equilibration buffer should be used in total for all three steps.



3 Sample application

- Add maximum 0.5 mL of sample to the column.
- For sample volumes less than 0.5 mL, add equilibration buffer to adjust the volume up to 0.5 mL after the sample has entered the packed bed completely.
- Let the sample or equilibration buffer enter the packed bed completely.
- Discard the flow-through.



4 Elution

- Place a test tube for sample collection under the column.
- Elute with 1.0 mL buffer and collect the eluate. A typical elution profile is shown in Fig 2


6 Spin protocol

1 PD MiniTrap G-25 preparation



- Remove the top cap and pour off the column storage solution.
- Remove the top filter using forceps.
- Remove the bottom cap.
- Put the PD MiniTrap G-25 into a 15 mL collection tube by using the column adapter, see Fig 1

2 Column equilibration



- Fill up the column with equilibration buffer and allow the equilibration buffer to enter the packed bed completely.
- Repeat once.
- Discard the flow-through.
- Fill up the column a third time with equilibration buffer and spin down at $1000 \times g$ for 2 minutes.
- Discard the flow-through.


Note: About 8 mL equilibration buffer should be used in total for all three steps.

3 Sample application



- Add sample (0.2–0.5 mL) slowly in the middle of the packed bed.

4 Elution



- Place the PD MiniTrap G-25 into a new 15 mL collection tube.
- Elute by centrifugation $1000 \times g$ for 2 minutes.
- Collect the eluate.

7 Recovery and desalting capacity

The following experiment is included as an example of a desalting experiment using the gravity protocol. A PD MiniTrap G-25 column was equilibrated with MilliQ™ water. 500 μL of bovine serum albumin 1000 $\mu\text{g}/\text{mL}$ in 1M NaCl was applied onto the column. The protein recovery was 95% and the desalting capacity was above 99%, see Fig 2."




Fig 2. Removal of NaCl from albumin solution with PD MiniTrap G-25. The albumin is eluted in volume fractions between 0.5 to 1.5 mL (indicated by arrows).

8 Column characteristics

Matrix	Sephadex G-25 resin
Particle size range	85 to 260 μm
Packed bed dimensions	0.97 \times 2.8 cm (2.1 mL)
Maximum sample volume	0.5 mL
Volume of eluted sample gravity	1.0 mL
Volume of eluted sample spin	0.5 mL
Desalting Capacity	>90%
Exclusion limit	M_r 5000
Chemical stability	All commonly used buffers
Working pH range	2–13
Storage temperature	+4 to - 30°C
Storage solution	0.15% Kathon CG/ICP biocide

9 Ordering information

Product	Pack size	Code No.
PD MiniTrap™ G-25	50	28-9180-07

Related products	Pack size	Code No.
PD-10 Desalting Columns	30	17-0851-01
PD SpinTrap G-25	50	28-9180-04
PD MultiTrap G-25	4 × 96-well filter plates	28-9180-06
PD MidiTrap G-25	50	28-9180-08
PD MiniTrap G-10	50	28-9180-10
PD MidiTrap G-10	50	28-9180-11
MiniSpin Adapter	10	28-9232-43
HiTrap™ Desalting	5 × 5 mL	17-1408-01
HiTrap Desalting ¹	100 × 5 mL	11-0003-29
HiPrep™ 26/10 Desalting	1 × 53 mL	17-5087-01
HiPrep 26/10 Desalting	4 × 53 mL	17-5087-02

¹ Pack size available by special order

www.gelifesciences.com/trap

GE Healthcare Bio-Sciences AB
Björkgatan 30
751 84 Uppsala
Sweden

GE monogram, SpinTrap, MultiTrap, MiniTrap, MidiTrap, HiTrap, HiPrep, and Sephadex are trademarks of GE Healthcare companies.

Kathon is a trademark of Dow Chemical Co., MilliQ is a trademark of Merck KGaA.

All third party trademarks are the property of their respective owners.

© 2007–2016 General Electric Company – All rights reserved.
First published Sep. 2007

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

GE Healthcare Europe GmbH
Munzinger Strasse 5,
D-79111 Freiburg,
Germany

GE Healthcare UK Ltd
Amersham Place
Little Chalfont
Buckinghamshire, HP7 9NA
UK

GE Healthcare Bio-Sciences Corp.,
100 Results Way
Marlborough
MA 01752
USA

GE Healthcare Dharmacon, Inc.,
2650 Crescent Dr,
Lafayette,
CO 80026
USA

HyClone Laboratories, Inc.,
925 W 1800 S, Logan
UT 84321
USA

GE Healthcare Japan Corporation
Sanken Bldg.
3-25-1, Hyakunincho
Shinjuku-ku, Tokyo 169-0073
Japan



28922529 AC 09/2016