### Instruction 22-0513-59 AC

Biacore™

# PDEA Thiol Coupling Reagent

## **Product description**


Order code:	BR-1000-58
Contents:	One bottle 2-(2-pyridinyldithio)ethaneamine hydrochloride, 100 mg
Storage:	+2 to 8°C
Safety:	For use and handling of the product in a safe way, please refer to the Safety Data Sheet.

Note: For in vitro use only.



### Intended use

PDEA Thiol Coupling Reagent is designed to be used for coupling biomolecules on Sensor Chip CM7, Sensor Chip CM5, Sensor Chip CM4, Sensor Chip CM3, and Sensor Chip C1 via thiol/disulphide exchange. Either the biomolecule or the sensor chip can be modified with PDEA to introduce reactive disulphide groups. Biomolecules are attached via introduced reactive disulphides or native thiol groups.



## Handling

For detailed information on immobilization strategies, including preparations of reagent solution and modification of ligand, refer to *Biacore Sensor Surface Handbook*, and to *LabGuides* on www.biacore.com/applicationsupporttools - Tech Tips & Protocols.

For local office contact information, visit www.gelifesciences.com/contact

GE Healthcare UK Limited Amersham Place Little Chalfont Buckinghamshire, HP7 9NA United Kingdom

www.gelifesciences.com/sampleprep

GE, imagination at work and GE monogram are trademarks of General Electric Company.

Biacore is a trademark of GE Healthcare companies.

© 2008-2013 General Electric Company – All rights reserved. First published June 2008

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

GE Healthcare Bio-Sciences AB Björkgatan 30, 751 84 Uppsala, Sweden

GE Healthcare Europe GmbH Munzinger Strasse 5, D-79111 Freiburg, Germany

GE Healthcare Bio-Sciences Corp. 800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA

GE Healthcare Japan Corporation Sanken Bldg. 3-25-1, Hyakunincho Shinjuku-ku, Tokyo 169-0073, Japan



22-0513-59 AC 08/2013