

Regeneration solutions

Product description

Buffer	Description	Package	Order code
Glycine 1.5	10 mM glycine-HCl pH 1.5	100 ml	BR-1003-54
Glycine 2.0	10 mM glycine-HCl pH 2.0	100 ml	BR-1003-55
Glycine 2.5	10 mM glycine-HCl pH 2.5	100 ml	BR-1003-56
Glycine 3.0	10 mM glycine-HCl pH 3.0	100 ml	BR-1003-57
NaOH 50	50 mM NaOH	100 ml	BR-1003-58

Note: All solutions are filtered through a 0.22 µm filter and ready to use.

Storage: Store at +2 to 8°C. The use-before date applies to solutions in unopened containers.

To maintain pH, the NaOH solution should be stored in plastic vials. The solution as provided may be aliquoted to plastic vials for storage.

Intended use: For removal of bound analyte from Biacore sensor surfaces.
Note: For *in vitro* use only.

Safety: For use and handling of the product in a safe way, please refer to the Safety Data Sheet.



For local office contact information, visit
www.gelifesciences.com/contact

GE Healthcare Bio-Sciences AB
Björkgatan 30
751 84 Uppsala
Sweden

www.gelifesciences.com/sampleprep

GE, imagination at work and GE monogram are trademarks of General Electric Company.

Biacore is a trademark of GE Healthcare companies.

© 1998-2013 General Electric Company – All rights reserved.
First published May 1998

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

GE Healthcare Europe GmbH
Munzinger Strasse 5, D-79111 Freiburg, Germany

GE Healthcare UK Limited
Amersham Place, Little Chalfont, Buckinghamshire, HP7 9NA, UK

GE Healthcare Bio-Sciences Corp.
800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA

GE Healthcare Japan Corporation
Sanken Bldg. 3-25-1, Hyakunincho Shinjuku-ku, Tokyo 169-0073, Japan



imagination at work

22-0567-48 AC 04/2013